

Calibration Certificate

Order Number

480607

Customer:

Wrights Backflow Testing & Services

Unit Under Test (UUT)

Manufacturer: Mid-West
Serial Number: 12190245
Model: 845-5
Gauge Range: 0-15 PSID

Calibration Information

Calibration Instrument: **DPI 610**
Serial Number: **61059948**
N.I.S.T. Traceable
Exceeds ASME B40.100 Section 6

Relative Humidity: 55

Temperature: 72

Performance Data:

<u>Calibration Point</u>	<u>Final Value</u>	<u>Deviation</u>
9 PSID	<u>9.1</u>	<u>0.1</u>
7 PSID	<u>7.1</u>	<u>0.1</u>
5 PSID	<u>5.0</u>	<u>0.0</u>
2 PSID	<u>1.9</u>	<u>-0.1</u>
1 PSID	<u>1.0</u>	<u>0.0</u>

Permissible Deviation: (+ or -) 0.2 PSID

We hereby certify that the differential pressure gauge referenced above has been calibrated to a (+ or -) 0.2 PSID of full scale accuracy on descending readings for the listed calibration points.

Certified By: Tom Jones

Date: 1/22/2021

American Backflow Products Co.

7580-A West Tennessee Street
Tallahassee, Florida 32304
850-575-9618 Fax:850-575-6508
www.backflowparts.com

Due Date: 1/22/2022



CERTIFICATION

Sunny Wright

Backflow Prevention Assembly Tester

Tester #: BFPACT0156

Expires: 1/1/2021

Course ID: 101635

Instructor: Mac Merritt

NC RWA Executive Director
PO Box 540 Wakefield, NC 27374
(336) 731-6903

Attach to Bid Attach to Bid Attach to Bid Attach to Bid Attach to Bid Attach to Bid

Addendum Acknowledgement Form

To: **BRUNSWICK COUNTY**

From: Bidder SUNNY WRIGHT

Address 658 BYRNES WAY SE
BOLIVIA NE 28422

Tele./Fax _____

Date of Bid 3-4-21

The undersigned, as bidder, hereby acknowledges receipt of Addenda numbered:

1 through 7

Respectfully submitted this day of 3-3-21

SUNNY WRIGHT BACKFLOW SERVICES

(Name of Firm or Corporation Making Bid)

By: Sunny Wright

Signature:

Name: Sunny Wright

Title: Owner

(Owner/Co-Partner/President/Vice President/Manager)